

103 - DIAPHRAGMATIC MUSCLE STRENGTHENING THROUGH FUNCTIONAL ELECTRICAL STIMULATION

VITOR TATSUO SAKAI
CESAR ANTONIO LUCHESA

Faculdade Assis Gurgacz - FAG , Cascavel , PR , Brazil
tatsuosakai@hotmail.com

INTRODUCTION

Currently respiratory therapy using a large arsenal of techniques to respiratory muscle strength, among these, are the breathing exercises, the use of equipment such as boosters and electrical stimulation. The latter provides a low cost, easy implementation and controversial results depending on the current literature.

Functional electrical stimulation (FES) is a form of treatment that uses low-frequency electrical current to cause contraction of muscles paralyzed or weakened due to upper motor neuron lesion , such as strokes, spinal cord injuries or cerebral skull , cerebral palsy , among other (DUARTE , Armenian FS , 2011) . The response to electrical stimulation of the diaphragm has been used over time for studying the mechanisms of action and as a means of ventilatory support in patients with high cervical spinal cord injury or in patients with alveolar hypoventilation. The principal clinical application of this technique, however, has been to evaluate the function of the phrenic nerve and diaphragm paralysis.

To evaluate the effect of muscle strengthening using an equipment which makes the measurement of inspiratory pressure (MIP - maximal inspiratory pressure, inspiratory FM) and expiratory (MEP - maximal expiratory pressure , expiratory FM) known for manometer . In this test, the volunteers performed three reproducible maneuvers, both the MIP , as for MEP , while the highest value was computed . All volunteers remained seated and making use of a nose clip to prevent escapes (NEDER, 1999) . The MIP has its normal value comprised in a young adult , in the range of -90 to -120 cmH₂O , whereas the MEP has its normal value comprised in a young adult , in the range of +100 to +150 cmH₂O . It is known that from the age of 20 is an annual decrease of 0.5 cmH₂O these values (Azeredo, 2002).

As for measuring the amount of air that enters and leaves the lungs spirometry, which can be accomplished with a slow breathing or forced expiratory maneuvers are used. The test helps in preventing and allows the diagnosis and quantification of respiratory disorders. Spirometry should be part of the assessment of patients with respiratory symptoms or known respiratory disease (Pereira , 2002) . For this test the subjects were stimulated vigorously to that carried out in an explosive exhalation maneuver onset and during exhalation the therapist observed the individual and encouraged positively to the effort is maintained for the time necessary. The objectives of this study are to analyze the manometer and spirometry before and after the application of FES on the diaphragm and compare respiratory conditions of patients after Technique in an attempt to prove the efficacy of functional electrical current to the respiratory muscle strengthening.

MATERIALS AND METHODS

This research is in a non- randomized clinical trial , conducted at the Rehabilitation Clinic FAG - Assisi School Gurgacz between the months of September- October 2013. The population was composed of individuals of both sexes , which were included in this study all those who met the criteria for inclusion and exclusion , and signed a consent form.

The sample consisted of 20 healthy individuals who contemplated the inclusion criteria of this study , the selection was made randomly. Being the inclusion criteria : age between 21 and 65 years, have any diseases of the respiratory system , with availability to attend the site on the date and time stipulated and signed the consent form . Exclusion criteria were : having pulmonary, cardiovascular or neurological diseases , present some contraindication to the device (FES) have any physical impairment that would prevent them from participating in the study , not accepting participate and / or do not sign the consent form under 21 years of age and with cognitive impairment . Was approved by the ethics committee of Assisi College Gurgacz with the number 156/2013 protocol.

Data collection was through an interview with the individual, individually in a room . The interview consisted of 6 questions , 3 open and 3 closed , followed by initial evaluation consisting of: Blood Pressure , Respiratory Rate , Heart Rate , Oximetry , ventilatory pattern , peripheral perfusion , Diaphragmatic Force (Scale Cuello) , Manovacuometry and Spirometry (Medical Microlab @ 3300) with the subject seated , with the application of FES parameters : pulse frequency : 40 Hz , pulse width : 250 ms , rise time : 1 second ; contraction time : 1 second ; fall time : one second ; relaxation time : 2 s ; stimulation time : 20 minutes; intensity : according to the sensitivity of the patient (Azeredo , 2002; Geddes , 1991) , in supine position and with arms along the body (was also asked who performed contractions synchronized with the firing of electrical current) and finally re : Manovacuometry Spirometry and again with the subject seated .

At the end of the collection , the present study was entirely typed and edited with Microsoft ® Word 2010 program. Data were tabulated and compared statistically with the IBM ® SPSS Statistics 20.0 software .

RESULTS AND DISCUSSION

Based on the survey it was found that the average age of individuals who participated in the survey ranges from 21 to 65 years. With an average age of 33.90 ± 12.37 years . As inferred the gender of the 20 individuals interviewed hears predominance in males, where sex: 70 % were male and 30 % female .

As with smoking , the prevalence of nonsmokers , which were found 55 % of non-smokers and 45 % smokers among nonsmokers some subjects had discontinued smoking for over ten years and it was found , so the literature says individuals who left the vice with more than 5 years time they run the same risk of cardiorespiratory diseases a person who has never smoked in his life . For averages of vital signs such as Respiratory Rate obtained 18.15 ± 2.23 , 74.60 ± 9.47 Heart Rate , $SpO_2 97.95 \pm 1.05$ and 60 % had a blood pressure of 120/80 mmHg, and can be seen that there was a good anthropometric homogeneity , since they are mostly within the normal range .

As shown in Table 01 all subjects had an increase in FEV₁ , FVC and FEV₁/FVC also the vast majority , thus proving the improvement in ventilatory capacity both inspiratory and expiratory in all subjects undergoing treatment with functional electrical current.

Sujeito	Pré-FES			Pós-FES		
	VEF1	CVF	VEF1/CVF	VEF1	CVF	VEF1/CVF
Sujeito 1	2,63	3,02	83	2,63	3,02	86
Sujeito 2	3,02	3,36	90	3,15	3,46	91
Sujeito 3	2,43	2,43	100	2,78	2,8	99
Sujeito 4	3,84	4,33	88	3,95	4,75	83
Sujeito 5	2,94	3,04	96	2,94	3,04	100
Sujeito 6	3,91	3,91	100	4,29	4,29	100
Sujeito 7	3,23	4,03	80	3,53	4,22	83
Sujeito 8	3,96	4,24	93	4,13	4,68	88
Sujeito 9	3,45	3,69	93	3,45	3,69	98
Sujeito 10	2,63	3,02	83	2,63	3,02	86
Sujeito 11	2,92	3,22	91	2,92	3,22	92
Sujeito 12	2,93	3,23	91	2,93	3,23	92
Sujeito 13	2,97	3,15	94	2,97	3,15	97
Sujeito 14	3,45	3,69	93	3,45	3,69	98
Sujeito 15	2,43	2,53	96	2,78	2,8	99
Sujeito 16	2,93	3,23	91	2,93	3,23	92
Sujeito 17	2,63	3,02	83	2,63	3,02	86
Sujeito 18	2,97	3,15	94	2,97	3,15	97
Sujeito 19	2,63	3,02	83	2,63	3,02	86
Sujeito 20	2,93	3,23	91	2,93	3,23	92
Média	3,04	3,33	90,65	3,13	3,44	92,25
Desvio padrão	0,45	0,50	5,60	0,49	0,58	5,79

Table 01 - Results and spirometry pre - post evaluation.

In Table 02 the values obtained through the manometer test, which found that all subjects in this trial had an increased respiratory muscle strength are presented. As much as the increase in inspiratory and expiratory pressure was not significantly elevated for all participants to be a single application of this protocol, the effects of the application of this technique have proven effective momentarily. According to Morgan et al., MIP and MEP have been used to identify the risk of respiratory failure and predict survival in patients with neuromuscular diseases.

Sujeito	Pré-FES		Pós-FES	
	PImax	PEmax	PImax	PEmax
Sujeito 1	80	100	100	120
Sujeito 2	80	80	90	90
Sujeito 3	120	90	150	100
Sujeito 4	140	120	170	160
Sujeito 5	100	90	110	110
Sujeito 6	140	100	180	140
Sujeito 7	120	120	140	160
Sujeito 8	130	130	160	150
Sujeito 9	80	80	100	100
Sujeito 10	90	115	120	130
Sujeito 11	80	80	110	100
Sujeito 12	110	120	140	160
Sujeito 13	120	110	150	130
Sujeito 14	110	100	130	120
Sujeito 15	90	80	115	120
Sujeito 16	130	130	140	150
Sujeito 17	140	120	160	140
Sujeito 18	120	120	140	130
Sujeito 19	110	110	130	140
Sujeito 20	130	120	150	150
Média	111,00	105,75	134,25	130,00
Desvio padrão	20,95	16,98	24,15	21,45

Table 02 - Results pre and post - manometer.

This research was limited to a single therapy session for each individual by the fact that most of the sample would be unavailable to stay back and continue the treatment. Another factor that contributed to the small sample size was that there were constant clashes between the hours of availability of subjects and the night called semesters. Therefore, we suggest to develop further studies with a larger quantity of sample and with a greater number of applications of this technique trying to keep track of activities conducted by sample and further study on this technique of improving respiratory capacity, which in this study was effective both in ventilation and in the degree of inspiratory and expiratory force.

CONCLUSION

We conclude this work, the application of FES was effective in increasing muscle strength, because there was an improvement in diaphragmatic activity in subjects who participated in this research. But having been a single application of this technique in each individual, gain muscle strength demonstrated in post FES review probably returned to baseline values, so further studies with a larger number of samples and technical applications can be realized.

REFERENCES

- AZEREDO, C.A.C. Fisioterapia Respiratória Moderna. 4. ed. São Paulo: Manole, 2002.
 AZEREDO, C.A.C. Fisioterapia Respiratória no Hospital Geral. São Paulo: Manole, 2000.
 BETHLEM, N. Pneumologia. 4. ed. São Paulo: Atheneu, 2002.
 CARVALHO, M. Fisioterapia Respiratória – Fundamentos e Contribuições. 5. ed. Rio de Janeiro: Revinter, 2001.
 COSTA, D. Fisioterapia Respiratória Básica. São Paulo: Atheneu, 1999.
 CRINER, G. et al. Variability of Electrophrenic Diaphragm Twitch Stimulation Over Time in Normal Subjects. Philadelphia: Respiratory Physiology, 1999.
 CUELLO, A.F.; MASCANTONIO, L.; MENDOZA, S.M. Estimulação Diafragmática Elétrica Transcutânea. [s.l.]: Medicina Intensiva, 1991.
 DUARTE, Armênio F. S.; VIEIRA, Mayla F. FES – Estimulação Elétrica Funcional. 2011. Disponível em < <http://www.fisioweb.com.br> >. Acessado em 19 de março de 2013.
 GEDDES LA, Voorhees WD, Babbs CF. Electroventilation, Proc. 5th Purdue Conf. CPR & Defibrillation. Am J Emerg

Med. 1985;3(4):337-9.

GEDDES LA, Voorhees WD, Lagler R, Riscili C, Foster K, Bourland JD. Electrically produced artificial ventilation. Med Instrument. 1988;22(5):263-71.

GEDDES LA, Simmons A. Artificial respiration in the dog by percutaneous bilateral phrenic nerve stimulation. Am J Emerg Med. 1991;9(6):527-9.

HAMIL, J.; KNUTZEN, K.M. Bases Biomecânicas do Movimento Humano. São Paulo: Manole, 1999.

JUNQUEIRA, L.C.; CARNEIRO, J. Histologia Básica. Rio de Janeiro: Guanabara Koogan, 1999.

KAPANDJI, A.I. Fisiologia Articular. 5. ed. São Paulo: Panamericana, 2000.

MORGAN, R.K., McNally, S., Alexander, M., Conroy, R., Hardiman, O., Costello, RW. Use of Sniff nasal-inspiratory force to predict survival in amyotrophic lateral sclerosis. Am J Respir Crit Care Med. 2005;171(3):269-74.

NEDER, JA, ANDREONI S, LERARIO MC, NERY LE. Reference values for lung function tests: II. Maximal respiratory pressures and voluntary ventilation. Braz J Med Biol Res. 1999;32(6):719-27.

PEREIRA, Carlos AC. Sociedade Brasileira de Pneumologia e Tisiologia. 2002. Disponível em <<http://www.sbpt.org.br/>>. Acessado em 04 de maio de 2013.

SCANLAN, C.; WILKINS, R.; STOLLER, J. Fundamentos da Terapia Respiratória de Egan. 7. ed. São Paulo: Manole, 2004.

SOUCHARD, P.E. O Diafragma. 2. ed. São Paulo: Summus Editorial, 1989.

TARANTINO, A.B. Doenças Pulmonares. 5. ed. Rio de Janeiro: Guanabara Koogan, 2002.

TAYLOR, AE. Clinical Respiratory Physiology. Saunders, 1989.

WEST, JB. Fisiologia respiratória moderna. 5th ed. São Paulo: Manole; 1996.

Rua: Pedro Ivo, 607 - Ap. 203.

CEP: 85812-171 Cascavel-PR

DIAPHRAGMATIC MUSCLE STRENGTHENING THROUGH FUNCTIONAL ELECTRICAL STIMULATION ABSTRACT

The aim of this study was to analyze the final results to prove the effectiveness of the application of functional electrical stimulation on diaphragm in increased respiratory muscle strength in healthy patients. The study included 20 healthy subjects of both sexes, with male predominance of 70% and with an average age of 33.90 ± 12.37 years. Initial evaluation, application of 20 minutes FES and final evaluation: a single session for each individual, approximately 40 minutes, where it was held. Where spirometry, maximal inspiratory pressure (MIP) and maximal expiratory pressure (MEP) were assessed before and after each session. Statistical analysis of data was performed by the Kolmogorov-Smirnov test ($p < 0.05$) except for initial FEV1/FVC ($p > 0.05$). All study subjects were positive and satisfactory results can therefore conclude that the application of FES was effective in increasing of respiratory muscle strength.

KEYWORDS: Electric Stimulation, Diaphragm, Muscular Strength.

RÉSUMÉ

Le but de cette étude était d'analyser les résultats définitifs pour prouver l'efficacité de l'application de la stimulation électrique fonctionnelle sur le diaphragme une augmentation de la force musculaire respiratoire chez les patients sains. L'étude a inclus 20 sujets sains des deux sexes, avec une prédominance masculine de 70% et avec un âge moyen de $33,90 \pm 12,37$ années. L'évaluation initiale, l'application de 20 minutes FES et évaluation finale: une seule session pour chaque individu, à environ 40 minutes, où il a eu lieu. Où spirométrie, pression maximale inspiratoire (MIP) et pression expiratoire maximale (MEP) ont été évalués avant et après chaque session. L'analyse statistique des données a été effectuée par le Kolmogorov-Smirnov test ($p < 0,05$) à l'exception de FEV1/FVC initiale ($p > 0,05$). Tous les sujets de l'étude ont été positifs et des résultats satisfaisants peuvent donc conclure que l'application de la FES a été efficace dans l'augmentation de la force musculaire.

MOTS-CLÉS: Stimulation Électrique, Diaphragme, Force Musculaire.

RESUMEN

El objetivo de este estudio fue analizar los resultados finales para probar la eficacia de la aplicación de la estimulación eléctrica funcional en el diafragma en un aumento de la fuerza muscular respiratoria en pacientes sanos. El estudio incluyó a 20 sujetos sanos de ambos sexos, con predominio de sexo masculino de 70% y con una edad media de $33,90 \pm 12,37$ años. Inicial de evaluación, aplicación de 20 minutos FES y evaluación final: se realizó una sola sesión para cada individuo, de aproximadamente 40 minutos, donde se llevó a cabo. Cuando la espirometría, la presión inspiratoria máxima (MIP) y la presión expiratoria máxima (PEM) fueron evaluados antes y después de cada sesión. El análisis estadístico de los datos se realizó mediante la prueba de Kolmogorov-Smirnov test ($p < 0,05$) a excepción de inicial FEV1/FVC ($p > 0,05$). Todos los sujetos de estudio fueron positivos y, por tanto, los resultados satisfactorios se puede concluir que la aplicación de FES fue eficaz en el aumento de la fuerza muscular.

PALABRAS CLAVE: Estimulación Eléctrica, Diafragma, Fuerza Muscular.

FORTALECIMENTO DA MUSCULATURA DIAFRAGMÁTICA ATRAVÉS DE ESTIMULAÇÃO ELÉTRICA FUNCIONAL

RESUMO

O intuito do presente estudo foi analisar os resultados finais a fim de comprovar a eficácia da aplicação de eletroestimulação funcional sobre diafragma no aumento da força muscular respiratória em pacientes saudáveis. Participaram do estudo 20 indivíduos saudáveis de ambos os sexos, sendo de domínio masculino de 70% e apresentando uma idade média de $33,90 \pm 12,37$ anos. Foi realizada uma única sessão para cada indivíduo, de aproximadamente 40 minutos, onde foi realizado: avaliação inicial, aplicação de 20 minutos de FES e reavaliação final. Foram avaliadas a espirometria onde, a pressão inspiratória máxima (PI_{max}) e pressão expiratória máxima (PE_{max}), antes e após cada sessão. A análise estatística dos dados foi realizada pelo teste Kolmogorov-Smirnov ($p < 0,05$) exceto para VEF1/CVF inicial ($p > 0,05$). Todos os sujeitos da pesquisa obtiveram resultados positivos e satisfatórios, podendo assim concluir que a aplicação de FES mostrou-se efetiva no aumento de força muscular respiratória.

PALAVRAS-CHAVE: Estimulação Eléctrica, Diafragma, Força Muscular.